



**Health**  
 Northern Sydney  
 Local Health District  
**SYDNEY DIALYSIS CENTRE**  
**HOME HAEMODIALYSIS MACHINE**  
**REQUIREMENTS**



FAMILY NAME		MRN
GIVEN NAME		<input type="checkbox"/> MALE <input type="checkbox"/> FEMALE
D.O.B.	/ /	M.O.
ADDRESS		
LOCATION / WARD		
COMPLETE ALL DETAILS OR AFFIX PATIENT LABEL HERE		

**Home haemodialysis equipment is installed under the following conditions:**

- All medical equipment remains the property Sydney Dialysis Centre (SDC). It is supplied on loan for the duration of the haemodialysis treatment. The equipment must be returned to SDC in the same condition as when installed.
- The treatment area will be large enough to accommodate all of the equipment with ample room for servicing (Approx. 2 square metres for the machine and RO) the room should also be free of clutter.
- Flooring shall be of water resistant quality and able to be cleaned of blood and chemical spills. Preferred flooring would be: Linoleum, Tiles, or Sealed timber.
- Carpet is not acceptable, as spills will ultimately occur and it will become damaged. SDC accepts no responsibility for damage incurred due to water or chemical spillage.
- The haemodialysis treatment area will be kept clean and free from pests and vermin at all times.
- Animals should not be allowed to enter the treatment area.
- Blood and chemical spills must be cleaned up as soon as possible to avoid contamination of the equipment and treatment area.
- Smoking is not permitted in the treatment area. All care must be taken to eliminate cigarette smoke from contaminating the equipment and minimum distance of 10 meters from a smoking area should be observed at all times.
- SDC technical staff must have access to attend the machine as required for the purpose of maintaining the equipment. The relationship between the technician, and all those present at the time of service, must be accommodating and respectful at all times.
- While every effort is made not to damage any property during installation, service and removal of equipment, SDC cannot accept responsibility for property damage caused during the course of these procedures.
- Technical staff are required to wear safety footwear at all times.
- The water, sewerage and electrical installation must comply with accepted standards as advised by SDC technical staff.
- Preferably the dialysis equipment should be located on the ground level of the residence. If a machine is required to be transported upstairs, an electrical stair walker must be able to be utilised.

***The above requirements adhere to WHS guidelines and as such SDC management can decide to remove the home dialysis equipment if the guidelines are not met & adhered to.***

I understand and accept this plan.

Signed in the presence of

Patient Name : \_\_\_\_\_

Staff Name : \_\_\_\_\_

Signature : \_\_\_\_\_

Signature : \_\_\_\_\_

Date : \_\_\_\_\_